

Tektronix (China) Co. Ltd.
1227 Chuan Qiao Road
Pudong New District, Shanghai, China 201206

TRACEABLE CALIBRATION
DPO2024B, C031252
Cal Date: 22-JUN-2016
Cal Due: Lab: Tek

Tektronix

Tektronix

CERTIFICATE OF TRACEABLE CALIBRATION

Certificate No: PC2KY63D5W

Revision: 00

Manufacturer: Tektronix

Model: DPO2024B

Serial No: C031252

Description: Oscilloscope; 200 MHz, 4 Channel

Temperature: 23.2 °C

Humidity: 44 %

Cal Date: 22-JUN-2016

Date Placed In Service: *

Due Date: *

*Optional customer entry fields: The "Due Date" may be established by adding the calibration interval to the "Date Placed In Service." Recommended calibration intervals, for Tektronix products, are available on the Tektronix website at:

http://www.tek.com/mytek_support_status/

INSTRUMENT CONDITION:

Received: Not applicable, initial calibration

Returned: In tolerance, within published measurement specification

Tektronix certifies the above instrument has been calibrated using standards traceable to the PRC National Institute of Metrology (NIM) and/or other National Metrology Institutes (NIST, NPL, PTB) that are linked to the international system of units (SI). The policies and procedures used for the calibration of this product comply with ISO/IEC 17025:2005.

This certificate shall not be reproduced, except in full, without the written approval of the calibration facility.

CALIBRATION PROCEDURE:

MANIFEST:Product_Oscilloscopes_DPO2000NGLF-2012_Full VERSION:57

CALIBRATION EQUIPMENT USED:

<u>MANUFACTURER/MODEL</u>	<u>MODEL DESCRIPTION</u>	<u>ID NUMBER</u>	<u>DUE DATE</u>
Fluke 9500B	Oscilloscope Calibrator	E300654	15-Oct-2016
Fluke 9530	Active Head, 9500	193387	15-Oct-2016
Fluke 9530	Active Head, 9500	CT0425	23-Jun-2017
Fluke 9530	Active Head, 9500	CT0479	15-Oct-2016
Fluke 9530	Active Head, 9500	CT0516	15-Oct-2016
Wavetek 9530	Active Head, 9500	187156	15-Oct-2016

Issued By:



Quality Director:

Cui, Hu Wa

Certified By: Chuanyong Li

Date Issued: 23-JUN-2016

ISO 9001 Registered Quality System

DPO2024B C031252

Page 1 of 1

001-1387-01

