

Tektronix (China) Co. Ltd.  
1227 Chuan Qiao Road  
Pudong New District, Shanghai, China 201206

TRACEABLE CALIBRATION  
TBS1102B, C014694  
Cal Date: 26-MAY-2016  
Cal Due: Lab: Tek

**Tektronix®**

**Tektronix®**

## CERTIFICATE OF TRACEABLE CALIBRATION

**Certificate No:** PC3KY3MQR4

**Revision:** 00

**Manufacturer:** Tektronix

**Model:** TBS1102B

**Serial No:** C014694

**Description:** Oscilloscope; 100MHz, 2 channel

**Temperature:** 25.4 °C

**Humidity:** 45 %

**Cal Date:** 26-MAY-2016

**Date Placed In Service:** \_\_\_\_\_ \*

**Due Date:** \_\_\_\_\_ \*

\*Optional customer entry fields: The "Due Date" may be established by adding the calibration interval to the "Date Placed In Service." Recommended calibration intervals, for Tektronix products, are available on the Tektronix website at:

[http://www.tek.com/mytek\\_support\\_status/](http://www.tek.com/mytek_support_status/)

### INSTRUMENT CONDITION:

**Received:** Not applicable, initial calibration

**Returned:** In tolerance, within published measurement specification

Tektronix certifies the above instrument has been calibrated using standards traceable to the PRC National Institute of Metrology (NIM) and/or other National Metrology Institutes (NIST, NPL, PTB) that are linked to the international system of units (SI). The policies and procedures used for the calibration of this product comply with ISO/IEC 17025:2005.

This certificate shall not be reproduced, except in full, without the written approval of the calibration facility.

### CALIBRATION PROCEDURE:

MANIFEST:Product\_Oscilloscopes\_TDSTBSTPS1K2K-9100-2012\_Full VERSION:84

### CALIBRATION EQUIPMENT USED:

<u>MANUFACTURER/MODEL</u>	<u>MODEL DESCRIPTION</u>	<u>ID NUMBER</u>	<u>DUE DATE</u>
Keithley 2000	Multimeter, 6 1/2 digit	CT0020	23-May-2017
Tektronix TBL-SGLF-9100	Table, Signal Generator Correction	SHOSLF03-01	14-Jun-2016
Tektronix TBL-SGLF-9100	Table, Signal Generator Correction	SHOSLF03-02	14-Jun-2016
Tektronix TBL-SGLF-9100	Table, Signal Generator Correction	SHOSLF03-03	14-Jun-2016
Tektronix TBL-SGLF-9100	Table, Signal Generator Correction	SHOSLF03-04	14-Jun-2016
Tektronix TBL-SGLF-9100	Table, Signal Generator Correction	SHOSLF03-05	14-Jun-2016
Wavetek 9100	Multifunction Calibrator	CT0017	20-Nov-2016

**Issued By:**



**Quality Director:** Cui, Hu Wa

**Certified By:** Zhengxing Wang

**Date Issued:** 27-MAY-2016

ISO 9001 Registered Quality System

TBS1102B C014694

Page 1 of 1

001-1387-01

