

Tektronix (China) Co. Ltd.
1227 Chuan Qiao Road
Pudong New District, Shanghai, China 201206



CERTIFICATE OF TRACEABLE CALIBRATION

Certificate No: PC8KY47YX4

Manufacturer: Tektronix

Description: Oscilloscope; 100MHz, 2 channel

Cal Date: 02-JUN-2016

Revision: 00

Model: TBS1102B

Temperature: 25.7 °C

Date Placed In Service: _____ *

Serial No: C014721

Humidity: 46 %

Due Date: _____ *

*Optional customer entry fields: The "Due Date" may be established by adding the calibration interval to the "Date Placed In Service." Recommended calibration intervals, for Tektronix products, are available on the Tektronix website at:
http://www.tek.com/mytek_support_status/

INSTRUMENT CONDITION:

Received: Not applicable, initial calibration

Returned: In tolerance, within published measurement specification

Tektronix certifies the above instrument has been calibrated using standards traceable to the PRC National Institute of Metrology (NIM) and/or other National Metrology Institutes (NIST, NPL, PTB) that are linked to the international system of units (SI). The policies and procedures used for the calibration of this product comply with ISO/IEC 17025:2005.

This certificate shall not be reproduced, except in full, without the written approval of the calibration facility.

CALIBRATION PROCEDURE:

MANIFEST:Product_Oscilloscopes_TDSTBSTPS1K2K-9100-2012_Full VERSION:84

CALIBRATION EQUIPMENT USED:

<u>MANUFACTURER/MODEL</u>	<u>MODEL DESCRIPTION</u>	<u>ID NUMBER</u>	<u>DUE DATE</u>
Fluke 9100	Multifunction Calibrator	CT0002	21-Jan-2017
Keithley 2000	Multimeter, 6 1/2 digit	KI10197	29-Aug-2016
Tektronix TBL-SGLF-9100	Table, Signal Generator Correction	SHOSLF18-01	14-Jul-2016
Tektronix TBL-SGLF-9100	Table, Signal Generator Correction	SHOSLF18-02	14-Jul-2016
Tektronix TBL-SGLF-9100	Table, Signal Generator Correction	SHOSLF18-03	14-Jul-2016
Tektronix TBL-SGLF-9100	Table, Signal Generator Correction	SHOSLF18-04	14-Jul-2016
Tektronix TBL-SGLF-9100	Table, Signal Generator Correction	SHOSLF18-05	14-Jul-2016

Issued By:

Quality Director:

Cui, Hu Wa

Certified By: Xinle Ge

Date Issued: 02-JUN-2016

ISO 9001 Registered Quality System

TBS1102B C014721

Page 1 of 1

001-1387-01

